

**EU DECLARATION OF CONFORMITY – No CETF006 rev1**

**Manufacturer:** Tunstall Healthcare (UK) Ltd  
Whitley Lodge, Whitley Bridge  
Yorkshire, DN14 0HR  
UK  
Tel: +44 1977 661234

This declaration is issued under the sole responsibility of the manufacturer.  
The manufacturer declares under our own responsibility that the following radio product(s):

Product name	Product number(s)	Product Description(s)
Gem+	P67005/15	PACKED-869 ALB GEM dark grey
	P67005/82	PACKED-869 AP GEM dark grey
	P67005/90	Packed Bogus Caller Gem

comply with the requirements of the following directive(s)


- 2014/53/EU Radio Equipment Directive (RED)**
- 2011/65/EU Restriction of the use of certain hazardous substances (RoHS)**


The conformity was checked in accordance with the following harmonised EN standard(s):

- EN 300 220-2:V3.1.1 (2017) EMC and Radio Spectrum Matters; Short Range Devices.**
- EN 55032:2012: Electromagnetic compatibility of multimedia equipment. Emission Requirements.**
- EN 301 489-3:V2.1.1 (2017) EMC standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD).**
- EN 50130-4: 2011 EMC Product Family Standard; Immunity Requirements for Components of Fire, Intruder and Social Alarm Systems.**
- EN 60950-1: 2006/A12:2011 Safety of Information Equipment.**

Signed

Date

  
\_\_\_\_\_  
Adam Miller

  
\_\_\_\_\_  
26/06/17

Position: Head of Devices & Platforms, Innovation and Development, Tunstall Healthcare (UK) Ltd