

## EU DECLARATION OF CONFORMITY – No CETF003 rev2

**Manufacturer:** Tunstall Healthcare (UK) Ltd  
 Whitley Lodge, Whitley Bridge  
 Yorkshire, DN14 0HR  
 UK  
 Tel: +44 1977 661234

This declaration is issued under the sole responsibility of the manufacturer.  
 The manufacturer declares under our own responsibility that the following radio product(s):

Part Number	Description
57215/810	Lifeline Smart Hub (CH)
57208/810	Lifeline Smart Hub (DE)
57205/810	Lifeline Smart Hub (EU)
57200/810	Lifeline Smart Hub (UK)
57208/622	Lifeline Smart Hub (MHD)

comply with the requirements of the following directive(s)

**2014/53/EU Radio Equipment Directive (RED)**  
**2011/65/EU Restriction of the use of certain hazardous substances (RoHS)**

The conformity was checked in accordance with the following harmonised EN standard(s):

Standard	Description
EN 300 220-2:V3.1.1 (2017)	EMC and Radio Spectrum Matters; Short Range Devices
EN 55032:2015	Electromagnetic compatibility of multimedia equipment. Emission Requirements
EN 50130-4: 2011 +A1:2014	Immunity Requirements for Components of Fire, Intruder and Social Alarm Systems
BS EN 61000-3-2:2006+A2:2009	Electromagnetic compatibility (EMC). Limits. Limits for harmonic current emissions (equipment input current $\leq 16$ A per phase).
BS EN 61000-3-3:2013 referencing (EN 301 489-7:V1.3.1, EN 301 489-3:V1.6.1, EN 301 489-1:V1.9.2 and EN 55024:2010 +A1:2015).	Electromagnetic compatibility (EMC). Limits. Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current $\leq 16$ A per phase and not subject to conditional connection.
EN 60950-1: 2006/A12:2011	Safety of Information Equipment

Signed

Date



Adam Miller

4/7/17

Position: Head of Devices & Platforms, Innovation and Development, Tunstall Healthcare (UK) Ltd